

# CERTIFICATE OF ANALYSIS No.: 2022-9520

## CLIENT

Natural Hemp Life, Panska 5 / 26  
00-124 Warszawa, Poland

## SAMPLE \*

Black Flamingo














Sample condition: SUITABLE  
Sample ID: 2231011  
Sample type: Plant material  
Batch No.: \*

Work order: 2022-106770  
Analysis ID: 2022\_175  
Method ID: PHL\_RPC\_12C  
Method SOP: MET-LAB-003-02

Sample received: 01/08/2022  
Start of analysis: 02/08/2022  
End of analysis: 03/08/2022  
Analyst: Janez Gerdenc

\* Information provided by the client.

CANNABINOID PROFILE		Concentration [% w/w]	Expanded uncertainty [% w/w]	Graphic presentation of relative cannabinoid concentration
<b>CBDV</b>	- Cannabidivarin	< LOQ	n/a	
<b>CBDA</b>	- Cannabidiolic acid	13.43	0.67	
<b>CBGA</b>	- Cannabigerolic acid	0.351	0.088	
<b>CBG</b>	- Cannabigerol	0.183	0.046	
<b>CBD</b>	- Cannabidiol	4.11	0.21	
<b>THCV</b>	- Tetrahydrocannavarin	0.098	0.021	
<b>CBN</b>	- Cannabinol	< LOQ	n/a	
<b>Δ<sup>9</sup>-THC</b>	- Δ-9-Tetrahydrocannabinol	0.19	0.073	
<b>Δ<sup>8</sup>-THC</b>	- Δ-8-Tetrahydrocannabinol	< LOQ	n/a	
<b>CBL</b>	- Cannabicyclol	0.071	0.016	
<b>CBC</b>	- Cannabichromene	0.283	0.048	
<b>Δ<sup>9</sup>-THCA</b>	- Δ-9-Tetrahydrocannabinolic acid	0.291	0.049	
<b>CBE</b>	- Cannabielsoin	< LOQ #	n/a	
<b>CBNV</b>	- Cannabivarin	< LOQ #	n/a	
<b>CBCA</b>	- Cannabichromenic acid	0.592 #	0.059	
<b>CBT</b>	- Cannabicitran	0.236 #	0.040	

Units and abbreviations: % w/w = weight percent, < LOQ = below the limit of quantitation (0.03 % w/w), ND = not detected, n/a = not available.


The results given herein apply only to the sample as received. Expanded Uncertainty was calculated using coverage factor k = 2, corresponding to a double standard uncertainty and characterizes the interval value in which it is possible to expect the real value with a probability of 95%. This is stated according to the ISO/IEC Guide 98-3.

Total or partial reproduction of this document is not allowed without the permit from PharmaHemp d.o.o. The document does not substitute any other legal document.

Date issued:

03/08/2022

Approved by:



mag. Marko Dragan  
Analytical Laboratory Manager

Authorized by:



dr. Boštjan Jančar  
Chief Technology Officer

End of Certificate